

Biofeedback Xpert System requirements

CF51208, July 2019

Computer

- > Processor: Intel i5 or AMD equivalent
- > Memory: 4 GB RAM
- > Graphics: Dedicated graphics card with 1 GB VRAM (not RAM)
- > Audio: Windows-compatible sound card with loudspeakers or headphones
- > Hard drive: 5 GB storage space
- > DVD drive
- > Mouse, keyboard
- > Monitor
- > One free USB port (at least USB 2.0) for the DONGLE
- > One free USB port (at least USB 2.0) for marker cable or control of an external device (only if used)
- > Operating system: Windows 7, Windows 8, Windows 10

Safety devices

If the Biofeedback is used in health care facilities the use of the following devices may be mandatory:

- > Isolating transformer for medical equipment according to EN 60601
- > Galvanic (medical) network isolation according to EN 60601 (if the computer is connected to a data network)

Please inquire with your company's safety representative.

Products of SCHUHFRIED GmbH are developed and in accordance with the requirements of the European Union guideline 93/42/EEG. The CE mark proves that safety-relevant regulations, EMC Standards for Medical Devices, Biocompatibility Evaluation of Medical Devices, product specific regulations and the underlying quality management system are adhered to.

Please contact us before acquiring new equipment so that we may give you the best possible advice.